

To: [5.1.5] [5.1.5 @who.int]; [5.1.2e] [5.1.2e @who.int]; [5.1.2e] [5.1.2e @rivm.nl];
 [5.1.2e] [5.1.2e @who.int]
Cc: [5.1.2e] [5.1.2e @rivm.nl]
From: [5.1.2e]
Sent: Mon 5/25/2020 8:56:22 PM
Subject: RE: [EXT] question from [5.1.2e]
Received: Mon 5/25/2020 8:56:54 PM

Dear [5.1.2e]

I see that your email below was sent few hours before my colleagues in Standards sent you your login details. Please let me know if you require any more assistance in ordering the research reagent and panel.

We provide the research reagent and panel in our catalogue to help laboratories worldwide developing serological assays. We have given some "representative results" in the datasheet, but we are not claiming to have a reagent to be used as calibrant.

We are working towards the establishment of the 1st WHO International Standard for anti-SARS-CoV-2 antibody and WHO Reference Panel. The International Collaborative study will start at the end of the June, with the aim of having the International standard available by January 2021. The WHO International Standards are the highest order of reference material, but as such the timelines for their establishment are relatively long. The current research reagent 20/130 will be included in the collaborative study and calibrate to the International Standard, so those using 20/130 in the meantime will be able to back calibrate their assays.

If you are interested in joining the WHO collaborative study, please let me know and I will be happy to send you the relative information.

Kind regards

[5.1.2e]

From: [5.1.5] <[5.1.5 @who.int]>

Sent: 25 May 2020 07:18

To: [5.1.2e] <[5.1.2e @who.int]>; [5.1.2e] <[5.1.2e @rivm.nl]>; [5.1.5] <[5.1.5 @who.int]>; [5.1.2e] <[5.1.2e @who.int]>

Cc: [5.1.2e] <[5.1.2e @rivm.nl]>; [5.1.2e] <[5.1.2e @nibsc.org]>

Subject: RE: [EXT] question from [5.1.2e]

Dear colleagues,

The validation panel provided by NIBSC is meant to validation assays across the world, not only in Europe.

There is still material if you wish to order, the sera panel and high titer research reagent 20/130, procedure attached. I am ccing my colleague [5.1.2e] that can you provide more info about the availability.

Best,

[5.1.2e]

[5.1.2e] [5.1.2e]

[5.1.2e] [5.1.2e]

[5.1.2e]

[5.1.2e]

[5.1.2e]

Dubbel

5.1.2e

5.1.2e

5.1.2e

5.1.2e

5.1.2e

5.1.2e

Dubbel

Dubbel