



Overview results

monster ID	Sample informatie		Correcte resultaten		
	pool sera IC patienten	verdunning	Test	%	n
1	SARS-CoV-2	1 (1:2?)	gehoeligheid	100	31/31
6	SARS-CoV-2	2 (1:3?)	gehoeligheid	100	32/32
7	SARS-CoV-2	3 (1:4?)	gehoeligheid	100	32/32
3	SARS-CoV-2	4 (1:10?)	gehoeligheid	97	30/31
4	SARS-CoV-2	5 (1:20?)	gehoeligheid	58	18/31
2	negative serum <2020	1,000	specificiteit	97	31/32
5	negative serum <2020	1,000	specificiteit	100	32/32



Each color is one lab -> seems to be a bit of a pattern, i.e. always under or above the median

- I saw no pattern for handmatig vs. automatisch
 - I also saw no pattern for the different platforms -> maybe 'Immunomat' performed a bit worse than DS2 and BEP (but not significant)
- if you want I can make some figures?