

Beoordeling Abena isolatiejassen

Documenten:

Mail FW Isolatiejassen – snel schakelen

Mail van dir. Abena aan LCH

- Partij 1: Isolation gowns PP 22gsm +PE12gsm, half laminated, light yellow. CAT I, PPE
- Partij 2: ISOLATION GOWN - STITCHING - ELASTIC CUFF - M - 115 X 140 CM; 33 GR/M2 PPPE (PP 18GSM / PE 15GSM) - FULLY LAMINATED - LIGHT BLUE. Documentatie: **The gowns are tested** according to below standards and reports can be send; EN 14126:2004; EN ISO 13688: 2013

Foto van partij 1 bijgevoegd, niet van partij 2.

Testrapporten zijn niet beschikbaar in map Abena.

Isolation gown foto

Foto van gele jas in plastic.

Opschrift ASID BONZ, CE keurmerk, LOTnr, prod/exp-data. Geen normen.

31032020 Opdracht Isolation Gowns partij 1 50000 stuks

Bijlage in mail van dir. Abena aan LCH.

In condities van de order staat:

- The product is covered by applicable parts of SS-EN-166 and related standards SS-EN-167 and SS-EN-168, but does not fulfill these standards. The product has not been tested or approved by officially recognized certification authorities.
- Abena does not issue a CE-mark, and it is buyer's responsibility to ensure and obtain permission from the authorities to use the product. The customer carries the risk for this himself.

Genoemde normen gaan over oogbescherming en zijn dus sowieso niet van toepassing.

Proforma factuur 31-03-2020

Afleveradres Vianen

Levering door Mediq aan LCH

CONCLUSIE: geen documenten beschikbaar om kwaliteit van geleverde product te beoordelen.

Volgens mail Abena zouden deze wel op te vragen zijn.