





中国认可 国际互认 检测 TESTING CNAS L0599

Test Report SL52025256570301TX Date:May 18,2020 Page 1 of 3

MAJESTIC SAFETY PRODUCTS & SERVICES B.V. P.O. BOX 80, 3200 AB SPIJKENISSE, THE NETHERLANDS.

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)Disposable surgical face mask: Type IIR (Claimed Type IIR)

Style No. : M-Safe 5053
Composition : (A)PP
Sample Color : (A)Blue
End Use : (A)health care

Buyer : EU Order No. : M12

Order No. : M123899
Country of Destination : The Netherlands

Country of Origin : China

Proposed Care Instruction: -

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : May 06, 2020

Testing Period : May 06, 2020 - May 18, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

#### Comment:

Commence	
Medical Face Masks-Requirements and Test Methods(EN 14683:2019)	(A)
Clause 5.2.2 Bacterial filtration efficiency (BFE)	M
Clause 5.2.3 Breathability	M
Clause 5.2.4 Splash Resistance	M
Clause 5.2.5 Microbial Cleanliness	M

Remark: M=Meet EN 14683:2019 Type IIR requirement

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

5.1.2e



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx.and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions.aspx.and, for electronic format documents. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document sunlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing finspection report & certificate, please contact us at telephone: (86-755)83011443,



Test Report

SL52025256570301TX

Date: May 18,2020

Page 2 of 3

Test Result

# Medical Face Masks-Requirements and Test Methods

(EN 14683:2019)

## Clause 5.2.2 Bacterial filtration efficiency (BFE)\*\*

1# 2# 3# 4# 5# (BFE), % >99.9 >99.9 >99.9 >99.9

Remark: Performance Requirement: Type I>95%, Type II>98%, Type IIR >98%

\*\*: This test standard is not within the accredited scope in SGS Shanghai testing centre, it is carried out by external laboratory accredited by CMA (China Metrology Accreditation)

## Clause 5.2.3 Breathability (Differential Pressure)

(EN 14683 :2019 Annex C, Flow rate 8 I/min)

	1#	2#	3#	4#	5#
Differential pressure △P (Pa/cm²)	37	38	37	38	38

Remark: Performance Requirement: Type I<40 Pa/cm<sup>2</sup>, Type II<40 Pa/cm<sup>2</sup>, Type IIR<60 Pa/cm<sup>2</sup>

### Clause 5.2.4 Splash Resistance

(ISO 22609 :2004, Pressure 16.0 kPa)

enetration	on inside sur	face		2-12			
1#	2#	3#	4#	5#	6#	7#	8#
Pass	Pass	Pass	Fail	Pass	Pass	Pass	Pass
9#	10#	11#	12#	13#	14#	15#	16#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
17#	18#	19#	20#	21#	22#	23#	24#
Pass	Pass	Pass	Pass	Fail	Pass	Pass	Pass
25#	26#	27#	28#	29#	30#	31#	32#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
Number of P	ass:		30				
Overall result:			Acceptable				

#### Domark

- 1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR: Y16.0kPa
- 2) Distance of the medical face mask target area surface to the tip of cannula is 300±10mm.
- 3) Condition and Test temperature (21±5)r C, relative humidity (85±10)%
- 4) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic Footments at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document satisfies divised that information contained hereon reflects the Company's findings at the time of its invervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exceptable the stransaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company, Any unauthorized alteration, forgery or faisification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the resource of the content or the surface of the content or the surface of the content or the surface of the surface of the content or the surface of the su

or email: 5.1.2e <u>@sgs.com</u> 3<sup>rd</sup>Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666

f (86-21) 64958763

www.sgsgroup.com.cn e 5 1 2e @sgs.com



Test Report SL52025256570301TX Date: May 18,2020 Page 3 of 3

Clause 5.2.5 Microbial Cleanliness

(EN 14683: 2019 Annex D)

1# 2# 3# 4# 5# CFU/g 14 22 21 18 13

Remark: Performance Requirement: Type ≤30 CFU/g, Type II≤30 CFU/g, Type IIR≤30 CFU/g



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com/en/Terms-and-Conditions.aspx.and">http://www.sgs.com/en/Terms-and-Conditions.aspx.and</a>, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx.Attention">http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx.Attention</a> is drawn to the limitation of liability, indemnification and jurisdiction issues defined rein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, foregory or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) itself and such sample(s) are retained for 30 days only.

or email: 200233 3<sup>®</sup>Bulding,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐江区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666

f (86-21) 64958763

www.sgsgroup.com.cn e 5.1.2e @sgs.com