



MNA LABORATORIES TEST REPORT

Report No : M-2020-00284	Date: 15.09.2020	Page: 1 / 2	Rev:
--------------------------	------------------	-------------	------

Purpose of Analysis : SPECIAL REQUEST	Brand : JEDX FFP3
Sample Type : PROTECTIVE MASK	Model :
Sample Send Org. : JEDX MEDCARE	Sampler : CUSTOMER
Manufacturer Name : JEDX MEDCARE	
Analysis Date : 09.09.2020	
Sample Quantity : 100 pieces	
Other informations : JEDX FFP3	

No	Tests	Results	Limit Value	Method	Evaluation	Physical Condition	
1	Total Inward Leakage (Mask)	General:1,1 Subjects:1,1 (%)	General≤5 Subjects≤2	TS EN 149+A1 Part 8.5	PASS		
2	Flammability (Mask)	No flame seen.	Max. 5 s flame	TS EN 149+A1 Part 8.6	PASS		
3	Carbon Dioxide Content of the Inhalation Air	0,75 (%)	Max%1	TS EN 149+A1 Part 8.7	PASS		
4	Exhalation Resistance	2,5 (mbar)	FFP3≤ 3,0 mbar	TS EN 149+A1 Part 8.9.2	PASS		
5	Inhalation Resistance	30 L/dk	0,4 (mbar)	FFP3≤ 1,0 mbar	TS EN 149+A1 Part 8.9.3		PASS
		95 L/dk	1,9 (mbar)	FFP3≤ 3,0 mbar	TS EN 149+A1 Part 8.9.3		
6	Penetration of Filter Material	Sodium Chloride	0,3 (%)	FFP3≤ 1 %	BS EN 13274-7	PASS	
		Paraffin Oil	0,2 (%)	FFP3≤ 1 %	BS EN 13274-7		

SAMPLE PLACE



MNA LABORATORIES TEST REPORT

Report No : M-2020-00284	Date: 15.09.2020	Page: 2 / 2	Rev:
Purpose of Analysis : SPECIAL REQUEST	Brand : JEDX FFP3		
Sample Type : PROTECTIVE MASK	Model :		
Sample Send Org. : JEDX MEDCARE	Sampler : CUSTOMER		
Manufacturer Name : JEDX MEDCARE			
Analysis Date : 09.09.2020			
Sample Quantity : 100 pieces			
Other informations : JEDX FFP3			

Operating as an experimental laboratory, MNA Laboratories have been accredited by TURKAK with AB-1183-T and TS_EN_ISO / IEC_17025: 2017 standard. Turkish Accreditation Agency (TÜRKAK) signed a multilateral agreement with the European Accreditation Association (EA) on the recognition of test reports and a mutual recognition agreement with the International Laboratory Accreditation Association (ILAC).

* Analysis is under accreditation.

Note :

1. No part of this analysis report can be used alone or separately, and may not be partially copied or reproduced, used to third parties and as a means of advertising without the written permission of the laboratory.
2. Analysis results are valid for the above mentioned sample sent by MNA Laboratory company / institution / person. It may not represent the whole.
3. Unsigned and unsealed reports are invalid.
4. This analysis report cannot be used in judicial-administrative procedures and for advertising purposes.
5. Results are valid for the sample as received.
6. The decision rule is the rule that determines how measurement uncertainty is taken into account when specifying the PASS density to a specified specification. According to the TLM-052 Decision Rule Implementation instruction, the Decision Rule Implementation Method selected in agreement with CUSTOMER is clearly stated in the report.
7. Limit Values are determined by taking from analysis methods.
8. The laboratory is not responsible if the information provided by the CUSTOMER affects the validity of the results.
9. Test and / or measurement results, expanded measurement uncertainties (if any) and test methods are given in the following pages, which are the supplementary part of this certificate.
10. Water Repellency Determination Hydrostatic Pressure Determination TS ISO 811 (Hydrostatic Pressure Tester E / N: 53) Analysis, Seam Strength EN ISO 13965-2 (Strength Test Device E / N: 50) Analysis and resistance to liquid chemical permeation TS EN 659 -A1 Part 3.18 (Liquid Chemical Transfer Device E / N: 107) Analysis is carried out in the conditioning room and ISO 139 PART 3.2 conditions (23 ± 2 ° C temperature and $50 \pm 4\%$ relative humidity) are applied for ambient conditions.
11. List of phthalates analyzed is below.
 Di-iso-nonyl phthalate (DINP), CAS number: 28553-12-0 or 68515-48-0
 Di- (2-ethylhexyl) phthalate (DEHP), CAS number: 117-81-7
 Di-n-octyl phthalate (DNOP), CAS number: 117-84-0
 Di-iso-decyl phthalate (DIDP), CAS number: 26761-40-0 or 68515-49-1
 Butyl benzyl phthalate (BBP), CAS number: 85-88-7
 Di-butyl phthalate (DBP), CAS number: 84-74-2

5.1.2e

5.1.2e

5.1.2e