

Glocal Linker, Healthcare Better; Dream Maker, Win Together >>>





Disposable Face Mask Product Profile

KA International Business Dept

Lyncmed Medical Group

Lyncmed Product Profile



3 Layer Non-woven Disposable Face Mask

Disposable Face Mask Key Features:

- Skin Friendly High Quality PP Material, 3-Ply
- Low Breathing Resistance, Bacterial Filtration Efficiency(BEF)>98%
- Ear Loop, Elastic Band, Latex Free
- Anatomic Adjustable Integrated nose bridge
- Size:17.5*9.5cm



Face Mask Packing Info



Packing Size: 52*38*30cm
Net Weight: 6.0KG
Gross Weight: 6.5KG
Quantity: 2000pcs/carton
(40box/carton,50pcs/box)



Face Mask Certificates –EN14683 BEF Test Report



Spencer
5.1.2e
Lyncmed Medical Technical (Beijing) Co., Ltd
Room 119, No. 1111
South Huihe Road, Chaoyang District
Beijing, 100000
CHINA

Bacterial Filtration Efficiency (BFE) and Differential Pressure (Delta P) Final Report

Test Article: Product Name: Non-woven Face mask
LOT No: CMA4714
Study Number: 108913-501
Study Received Date: 23 Aug 2018
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP004 Rev 15
Deviation(s): None

Summary: The BFE test is performed to determine the filtration efficiency of test articles by comparing the bacterial control counts upstream of the test article to the bacterial counts downstream. A suspension of *Staphylococcus aureus* was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at $1.7 - 2.7 \times 10^7$ colony forming units (CFU) with a mean particle size (MPS) of $3.0 \pm 0.3 \mu\text{m}$. The aerosols were drawn through a six-stage viable particle, Andersen sampler for collection. This test method complies with ASTM F2101-14, EN 14683:2014, Annex B, and AS4381:2015.

The Delta P test is performed to determine the breathability of test articles by measuring the differential air pressure on either side of the test article using a manometer, at a constant flow rate. The Delta P test was designed to comply with MIL-M-36954C, Section 4.4.1.2 and complies with EN 14683:2014, Annex C and AS4381:2015.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Inside
BFE Test Area: $\sim 40 \text{ cm}^2$
BFE Flow Rate: 28.3 Liters per minute (L/min)
Delta P Flow Rate: 8 L/min
Conditioning Parameters: $85 \pm 5\%$ relative humidity (RH) and $21 \pm 5^\circ\text{C}$ for a minimum of 4 hours
Test Article Dimensions: $\sim 177 \text{ mm} \times \sim 158 \text{ mm}$
Positive Control Average: 2.5×10^7 CFU
Negative Monitor Count: <1 CFU
MPS: $3.1 \mu\text{m}$



10 Sep 2018
Study Completion Date

5.1.2e



801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com
Page 1 of 2



Study Number: 108913-501
Bacterial Filtration Efficiency (BFE)
and Differential Pressure (Delta P) Final Report

Results:

Test Article Number	Percent BFE (%)	Delta P (mm H ₂ O/oz)	Delta P (Pa/cm ²)
1	99.8	3.6	35.2
2	99.9	3.6	35.6
3	99.7	3.7	35.9
4	>99.9*	3.4	33.5
5	99.9	3.8	36.8

* There were no detected colonies on any of the Andersen sampler plates for this test article.

The filtration efficiency percentages were calculated using the following equation:

$$\% \text{ BFE} = \frac{C - T}{C} \times 100$$
 C = Positive control average
 T = Plate count total recovered downstream of the test article
 Note: The plate count total is available upon request

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com
Page 2 of 2

Face Mask Certificates –EN14683 MC Test Report



Sponsor:
Mavis CUI
Lynomed Medical Technical (Beijing) Co.
Room 119, No. 1111, South Huihe Rd., Chaoyang District
Beijing, 100000
CHINA

Microbial Cleanliness (Bioburden) of Medical Masks Final Report

Test Article: Product Name: Non-woven Face mask
LOT #CMAA714
Study Number: 1088914-001
Study Received Date: 23 Aug 2018
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0036 Rev 14
Customer Specification Sheet (CSS) Number: Z01805306 Rev 01
Deviation(s): None

Summary: The testing was conducted in accordance with EN 14683:2014, with the exception of approximate volumes of eluent used when performing the extraction procedure and a temperature range of 30-35°C used for aerobic incubation.

When bioburden results are calculated using a software program, manual calculations may differ slightly due to rounding. The counts determined on products are colony forming units and may not always reflect individual microorganisms. The sponsor performs any statistical analysis and determines the acceptable limits. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Results:

Unit Number	Weight (g)	Aerobic	Fungal	Total Bioburden (CFU/mask)	Total Bioburden (CFU/g)
1	3.2	88	12 ^a	99.7	31.1
2	3.3	49	3 ^a	51.9	15.7
3	3.3	33	3 ^a	35.9	10.9
4	3.4	51	<3	54.2	15.9
5	3.4	88	9 ^a	77.5	22.8
Recovery Efficiency:		65.7%			

< = No Organisms Detected
Note: The results are reported as colony forming units (CFU) per mask.
Note: Sample positive testing was performed using *Bacillus atrophaeus*. The test article was not inhibitory using this test method.
^a Spreader. Count is considered a minimum estimate due to swarming of certain colonies on the membrane.

5.1.2e



Study Number 1088914-001
Microbial Cleanliness (Bioburden) of Medical Masks Final Report
A Sotera Health company

Test Method Acceptance Criteria: If applicable, anaerobic controls are acceptable for the bioburden test results. The number of masks to be tested shall be a minimum of 5 or more to meet an acceptable quality level of 4%. The bioburden of the medical mask shall be < 30 cfu/g tested.

Procedure:

Positive Controls/Monitors: *Bacillus atrophaeus*
Extract Fluid: Peptone Tween[®] with Sodium Chloride
Extract Fluid Volume: ~300 mL
Extract Method: Orbital Shaking for 5 minutes at 250 rpm
Plating Method: Membrane Filtration
Agar Medium: Tryptic Soy Agar
Sabouraud Dextrose Agar with Chloramphenicol
Recovery Efficiency: Exhaustive Rinse Method
Aerobic Bacteria: Plates were incubated 3 days at 30-35°C, then enumerated.
Fungal: Plates were incubated 7 days at 20-25°C, then enumerated.

Face Mask Certificates –EN14683 SBPR & EN14683 Blood Penetration Resistance Report



5.1.2e
LyncMed Medical Technical (Beijing) Co., Ltd.
Room 119, No. 1111,
South Huilhe Road, Chaoyang District
Beijing, 100020
CHINA

Synthetic Blood Penetration Resistance Final Report

Test Article: Product Name: Non-woven Face mask
LOT #CMA4714
Study Number: 1088912-501
Study Received Date: 23 Aug 2018
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0012 Rev 08
Deviation(s): None

Summary: This procedure was performed to evaluate surgical facemasks and other types of protective clothing materials designed to protect against fluid penetration. The purpose of this procedure is to simulate an arterial spray and evaluate the effectiveness of the test article in protecting the user from possible exposure to blood and other body fluids. The distance from the target area surface to the tip of the cannula is 30.5 cm. A test volume of 2 mL of synthetic blood was employed using the targeting plate method.

This test method was designed to comply with ASTM F1862 and ISO 22609 (as referenced in EN 14683:2014 and AS4381:2015) with the following exception: ISO 22609 requires testing to be performed in an environment with a temperature of 21 ± 5°C and a relative humidity of 85 ± 10%. Instead, testing was performed at ambient conditions within one minute of removal from the environmental chamber held at those parameters.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Number of Test Articles Tested: 32
Number of Test Articles Passed: 30
Test Side: Outside
Pre-Conditioning: Minimum of 4 hours at 21 ± 5°C and 85 ± 5% relative humidity (RH)
Test Conditions: 18.8°C and 32% RH

Results: Per ASTM F1862 and ISO 22609, an acceptable quality limit of 4.0% is met for a normal single sampling plan when 29 of 32 test articles show passing results.

Test Pressure: 120 mmHg (16.0 kPa)

Test Article Number	Synthetic Blood Penetration
1-18, 20-26, 28-32	None Seen
19, 27	Yes

5.1.2e

10 Sep 2018
Study Completion Date



1088912-501
NELSON LABS | nelsonlabs.com | sales@nelsonlabs.com

FORM: FRT0012-000 Rev. 10
Page 1 of 1

These results relate only to the test article listed in this report. Reports may not be reproduced except in their entirety. Subject to NLS Terms and Conditions at www.nelsonlabs.com



5.1.2e
Sponsor
LyncMed Group
No. 1111, South Huilhe Road,
100025, Beijing CHINA

EN 14683:2005 Synthetic Blood Penetration Resistance Final Report

Test Article: Non-woven Facemask
Laboratory Number: 735314
Study Received Date: 27 Jan 2018
Test Procedure(s): Standard Test Protocol (STP) Number STP0012 Rev 05

Summary: This procedure was performed to evaluate surgical facemasks and other types of protective clothing materials designed to protect against fluid penetration. The purpose of this procedure is to simulate an arterial spray and evaluate the effectiveness of the material in protecting the user from possible exposure to blood and other body fluids. The distance from the target area surface to the tip of the cannula is 30.5 cm. A test volume of 2 mL of synthetic blood was employed. This test method was designed to comply with ASTM F1862 and EN 14683:2005. All test method acceptance criteria were met.

Number of Test Articles Tested: 32
Number of Test Articles Passed: 32
Test Side: Outside
Pre-Conditioning: Minimum of 4 hours at 21 ± 5°C and 85 ± 5% relative humidity (RH)
Test Conditions: 22.5°C and 21% RH


5.1.2e

07 Feb 2014
Study Completion Date

FRT0012-000 Rev. 10
NELSON LABS | nelsonlabs.com | sales@nelsonlabs.com

FORM: FRT0012-000 Rev. 10
Page 1 of 1

These results relate only to the test article listed in this report. Reports may not be reproduced except in their entirety. Subject to NLS Terms and Conditions at www.nelsonlabs.com



Face Mask CE Certificate

Technical Data Sheet

Lynomed Medical Technology(Beijing) Co., Ltd.
EC DECLARATION OF CONFORMITY

Manufacturer: Lynomed Medical Technology(Beijing) Co., Ltd.
Room 119, No.1111 South Huihe Rd, Chaoyang District 100022 Beijing, China

whose single Authorized Representative: SUNGO Certification Company Limited
RM101, MAPLE HOUSE, 118 HIGH STREET, PURLEY, LONDON, ENGLAND

We declare under our sole responsibility that:

the medical device -Non-woven products:Surgical gowns, Surgical drapes, Surgical packs, Pouches, Caps, Coveralls, Facemasks, Suits, Underpad, Gloves, Towels, Wipes, Sheets, Packs, Shoe covers, Aprons, Gowns, Trousers, Sleeve covers, Pillow covers, Beard covers, Bags, Bibs, Rolls, Briefs, Hair bands;
-Disposable PE products: Shoe covers, Tray covers, Pouches, Caps, Gloves, Aprons, Gowns, Sleeve covers, Bags, Sheets, Pillow case, Pallet covers, Bibs;
-Disposable Vinyl examination gloves
-Disposable Nitrile examination gloves
-Disposable Latex examination gloves
-Urine Cup for single use
-Urine Bag for single use

(Name, type or model, batch or serial number, possibly sources and number of items)

of class **Class I**
according to annex IX of direct. 93/42/EEC

meets all the provisions of the directive 93/42/EEC (or 90/385/EEC) which apply to it.

Conformity assessment procedure 93/42/EEC, Annex V

The above mentioned declaration of conformity is exclusively under the responsibility of Lynomed Medical Technology(Beijing) Co., Ltd.

Place, date **5.1.2e** Management Representative
Beijing 20th Dec. 2017 Legally binding signature, function

5.1.2e

EC Declaration of Conformity

CE Certificate

TECHNICAL DATA SHEET - CONFIDENTIAL

This information shall not be disclosed to any other party

Product Medical Face Masks, 3 ply

Quality Norm EN 14683, Type I
BFE % \geq 95
Differential Pressure (Pa/cm²) \leq 29.4
Splash Resistance Pressure (mm Hg) not required
CIU/g < 30



Lab test lab test must be provided, showing that the lot meets the EN 14683 Type I requirements

Material Latex free, 3 ply mask:
1st ply, outside 25 gsm non-woven
2nd ply, 25 gsm melt blown filter
3rd ply, inside 25 gsm nonwoven

Size & fixture earloop: body 17,5 x 9,5 cm, ultrasonically welded latex free earloops at the outside of mask body

Packaging
primary 50 pcs per box with lot numer and color marked on the square on the box
secondary earloop case 40x50 pcs case white with full color label, size 52x38x30 cm

Intended Use The intended use of a medical face mask is to help prevent large particles expelled by the wearer (e.g. spit, mucus) from reaching the patient or work environment. Fluids contacting the outer surface of the surgical mask will not immediately soak through to the interior of the surgical mask and contact the wearer's lips or skin.

Wear Medical face masks are typically donned for a specific procedure. For infection control purposes masks are typically disposed of after each procedure/patient activity.

Fit Medical face masks are not designed to pass a fit test. Flat surgical masks fit loosely over the face leaving large gaps between the mask and the wearer. It is unlikely that most of the air will pass through the mask material. The air (and any airborne particles) will go through the gaps

Shelf life 5 years after production date


Lot to be traced back by the factory in case of recall or QA issues

QA documents MDD 93/42/EEC Guideline, EN 14683

Lyncmed Company Certificate



Business License



Export Customs License

Lyncmed Company Certificate



Our Bank Reputation Certificate



Medical Equipment Business Registration License

Lyncmed Company Certificate


TÜVRheinland

Certificate
The Certification Body of
TÜV Rheinland LGA Products GmbH

herby certifies that the organization
**Lyncmed Medical Technology
(Beijing) Co., Ltd.,
Room 119, Floor 1
GUOTOUSHANGKE Building
No. 1111, South Huihe Road
ChaoYang District
100022 Beijing
China**

has established and applies a quality management system for medical devices
for the following scope:
**Manufacture and Distribution of Medical Devices
(see attachment for products included)**

Proof has been furnished that the requirements specified in
EN ISO 13485:2016
are fulfilled. The quality management system is subject to yearly surveillance.

Effective Date: 2018-12-04
Certificate Registration No.: SX 60133386 0001
An audit was performed. Report No.: 50180180 002
This Certificate is valid until: 2021-12-03


Certification Body

5.1.2e


Deutsche
Akreditierungsstelle
D-24149-05-02

Date: 2018-12-04

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
Tele: +49 201 609-1211 Fax: +49 201 609-9020 e-mail: cert@cert.tuev-rheinland.com http://www.tuev-rheinland.com


TÜVRheinland


TÜV Rheinland
LGA Products GmbH
Tillystraße 2, 90431 Nürnberg

Doc. 1/1, Rev. 0

Attachment to
Certificate
Registration No.: SX 60133386 0001
Report No.: 50180180 002

Organization: **Lyncmed Medical Technology
(Beijing) Co., Ltd.
Room 119, Floor 1
GUOTOUSHANGKE Building
No. 1111, South Huihe Road
ChaoYang District
100022 Beijing
China**

Scope: Products:
Sterile Latex Surgical Gloves, Non-sterile Examination
Latex gloves, Dental High Speed Air Turbine Hand-
piece, Dental Low Speed Air Turbine Hand-piece,
Patient Examination Gloves, Ultrasonic scaler,
Radio motors, Apex locators, Pulp tester, Mask,
Laryngeal Mask Airway, Overall, Isolation Gown,
Sterile Surgical Gowns, Sterile Surgical Kite, Sterile
Surgical Caps, Sterile face masks, Sterile shoe covers,
Alcohol swabs, Diapers and Adult Diapers, Disposable
Under Pad, Medical Tape, Urine bag for Single Use,
Foley Catheter Silicone, Dental Alginate Wipers,
Vinyl Examination Gloves, Sterile Gauze Sponges

Certification Body

5.1.2e


Deutsche
Akreditierungsstelle
D-24149-05-02

Date: 2018-12-04


ISO-13485 Certificate



Product Workshop



Our Global Distribution Network



We make better healthcare accessible affordable and easily available

Our Warehouse



Warehouse Inside



Warehouse Outside



Warehouse Truck Loading

Thanks for all

Glocal Linker, Healthcare Better; Dream Maker, Win Together >>>