



Jiangyin Shine Color Textile Product Technology Service Co., LTD

Test Report

SC20200078

Page 1 of 2

Product Name	Disposable Protective Mask	Specification Type	FFP3 MASK
		Trademark	---
Trust Unit	FANG RUI	Tel	---
Manufacturer	---	Sample Grade	---
Sample Quantity	10	Sample Receiving Date	2020-05-12
Test Category	Entrusted inspection	Serial Number	051001
Samples Conditions	Meet the testing requirements		
Document and Decide Accordance	EN 149:2001+A1: 2009 《Respiratory protective devices-Filtering half masks to protect against particles-Requirements,testing,marking》		
Test Conclusion	The test results are shown on page 2		Signatuim Date:2020-05-14
Remarks	The sample was not pretreated. The company has not obtained qualification recognition, and the test results are only used for scientific research, teaching or internal quality control.		

Jiangyin qingzhicai textile products technical service co. LTD Address: no. 68-1 changle road, changshou, jiangyin city, jiangsu province Zip code: 214424

Telephone: 0510-5.1.2e

Note: 1. If there is any objection to the inspection report, it shall be submitted to the testing unit within 15 days from the date of receiving the report, and it will not be accepted after the deadline. 2. The report is invalid without signature or alteration, and the report is invalid without "special seal for testing".

Testing Results

SC20200078

Page 2 of 2

Serial	Test Items	Unit	Requirement	Results	Individual Judgment		
1	Penetration Of filter material	%	FFP1: ≤ 20 FFP2: ≤ 6 FFP3: ≤ 1	Sodium chloride test(95L/min)		---	
				Without pretreatment	1#		0.33
					2#		0.12
					3#		0.88
					4#		0.59
					5#		0.68
					6#		1.87
					7#		3.41
					8#		1.93
					9#		0.66
					10#		3.34

The product image



—————The following blank—————

Jiangyin qingzhicai textile products technical service co. LTD Address: no. 68-1 changle road, changshou, jiangyin city, jiangsu province Zip code: 214424 Telephone: 0510- 5.1.2e

Note: 1. If there is any objection to the inspection report, it shall be submitted to the testing unit within 15 days from the date of receiving the report, and it will not be accepted after the deadline. 2. The report is invalid without signature or alteration, and the report is invalid without "special seal for testing".