

Particle Penetration by Kalibra

Datum	01 mei 2020
Inkoopnummer	
LOT/X-nummer	X133

Q1 Particle Penetration	Gemiddelde rendement > 0,3	Pass or Fail t.o.v. P1/P2/P3	Gemiddelde rendement > 0,5	Pass or Fail t.o.v. P1/P2/P3
Masker 1	95%	Pass P2	98%	Pass P2
Masker 2	95%	Pass P2	99%	Pass P3
Masker 3	95%	Pass P2	98%	Pass P2
Totaal gemiddelde	95%	Pass P2	98%	Pass P2

Ademweerstand:

	Gemeten (< Pa)	Eis (< Pa)	Conclusie
Masker 1	88	< 240 Pa	Pass
Masker 2	116		Pass
Masker 3	106		Pass

Eindoordeel > 0,3
Voldoet P2
Eindoordeel > 0,5
Voldoet P2