

Particle Penetration by Kalibra

Datum	01-05-2020
Inkoopnummer	
LOT/X-nummer	X134 – FIOD-04

Q1 Particle Penetration	Gemiddelde rendement > 0,3	Pass or Fail t.o.v. P1/P2/P3	Gemiddelde rendement > 0,5	Pass or Fail t.o.v. P1/P2/P3
Masker 1	99 %	Pass P3	99 %	Pass P3
Masker 2	95 %	Pass P2	98 %	Pass P2
Masker 3	96 %	Pass P2	99 %	Pass P3
Totaal gemiddelde	97 %	Pass P2	99 %	Pass P3

Ademweerstand:

	Gemeten (< Pa)	Eis (< Pa)	Conclusie
Masker 1	155	< 240 Pa	Pass
Masker 2	100		Pass
Masker 3	117		Pass

Eindoordeel > 0,3
Voldoet aan P2 filterkwaliteit
Eindoordeel > 0,5
Voldoet aan P3 filterkwaliteit