

Particle Penetration by Kalibra

Datum	01 mei 2020
Inkoopnummer	
LOT/X-nummer	X135

Q1 Particle Penetration	Gemiddelde rendement > 0,3	Pass or Fail t.o.v. P1/P2/P3	Gemiddelde rendement > 0,5	Pass or Fail t.o.v. P1/P2/P3
Masker 1	96%	Pass P2	98%	Pass P2
Masker 2	93%	Pass P1	97%	Pass P2
Masker 3	93%	Pass P1	97%	Pass P2
Totaal gemiddelde	94%	Pass P2	97%	Pass P2

Ademweerstand:

	Gemeten (< Pa)	Eis (< Pa)	Conclusie
Masker 1	122	< 240 Pa	pass
Masker 2	87		pass
Masker 3	84		pass

Eindoordeel > 0,3
Voldoet P2
Eindoordeel > 0,5
Voldoet P2