

Eindbeoordelingsformulier mondkmaskers

Chirurgisch masker IIR

Lotnummer	1720
Leverancier/Terugkoppelen aan	
Productnaam of naam fabrikant	Yongli
Type mondkmasker (volgens de fabrikant)	974

AH/RIVM	Spattest	BFE
Datum:	Datum:	Datum:
☐ Pass ☐ Fail ☐ On Hold ☒ Spattest ☒ BFE	☒ Pass ☐ Fail ☐ On Hold	☒ Pass ☐ Fail ☐ On Hold
Uitleg indien On Hold:	Uitleg indien On Hold:	Uitleg indien On Hold:

Q3 en Q4	+	+/-	-
Oordeel RIVM			-
Oordeel AH/TOXI		+/-	
Algemene bevindingen: ALLE BEUGELS, NEUS METAALKERM			

Eindbeoordeling

Datum:

Artikelnummer VWS	Quality Check:
☒ 974 (Earloops) ☐ 984 (Lint)	☐ Goedgekeurd ☒ Afkeur
Algemene bevindingen: Afkeur omdat documentatie niet voldoet en wisselen de resultaten met spattest in verleden. Dit lot presenteert goed op spattest.	

Disclaimer: Deze rapportage bevat een onafhankelijk indicatief oordeel van de kwaliteit van de producten. Dit betekent dat deze resultaten slechts een gedeeltelijke weergave zijn van de kwaliteit van de producten. De rapportage is alleen te gebruiken door het LCH en het ministerie VWS ter ondersteuning bij het vrijgeven van producten op de markt. Gegevens uit deze rapportage mogen niet gedeeld worden met derden. Indien u niet de geadresseerde bent of dit bericht abusievelijk aan u is verzonden, wordt u verzocht dat aan de afzender te melden en de rapportage te verwijderen. Het LCH en het ministerie VWS aanvaarden geen aansprakelijkheid voor schade, van welke aard ook, die verband houdt met risico's verbonden aan het elektronisch verzenden van berichten.

Checklist prestatiecontrole adembeschermingsmaskers

Inkoopnummer		LOT-nummer	1770
Datum uitvoering check	3-7-2020 (Zita) 3-7-2020 (Zita)		
Leveringsdatum			
Leverancier	Yongli		
Productnaam of naam fabrikant			
Productiedatum			
Naam	5.1.2e		
Functie	5.1.2e		
Besproken met RIVM			
Advies: Type masker bij uitgifte:			

Werkwijze:

RIVM voert de "administratieve" controle uit met betrekking tot notified bodies, certificaten en Declarations of Conformities. De arbeidshygiënist beoordeelt opdruk binnendoos, verpakking masker op onregelmatigheden met behulp van deze checklist. Het masker zelf wordt gecontroleerd op aantal lagen, kwaliteit, bevestiging van elastieken en fit op het gezicht. Vervolgens wordt een gezamenlijk advies gegeven over onder welke categorie dit masker zal worden uitgegeven naar de ziekenhuizen.

Administratief (wordt ingevuld door RIVM)

Technical data sheet aanwezig?	☐ Ja	☐ Nee
Garantiecertificaat aanwezig?	☐ Ja	☐ Nee
Declaration of conformity aanwezig?	☐ Ja	☐ Nee
Oordeel van RIVM	<i>(nummer)</i> NoBo ontbreekt check levert op dat NoBo niet voor dit soort maskers mag worden gebruikt. Spattest door Nelson uitgevoerd → Spattest fail.	

Prestatie levering

Oordeel arbeidshygiënist	Voorheen waren verschillen in Lot 1642, 1770, 1769, 1638, 1640, 1768, 1639, 1641. (plastic neusbeugel in 1760) + In dit lot <u>alleen</u> metalen neusbeugels + Goede spattest, goede BFE, dit LOT NB. Spattest door Nelson uitgevoerd gaf ook fails, ook deze is toegestaan door leverancier. Guarany AH is ook dat spattesten soms fail zijn.
--------------------------	---

BINNENDOOS/PLASTICZAK	
1	Aantal maskers per binnendoos/plasticzak? 50
2	Algemene indruk binnendoos/plasticzak? (vochtplekken, ingedeukt, verkleurd etc) ☒ Integriteit in orde ☐ Integriteit niet in orde, omdat: ☐ Extra stickers opgeplakt, nl:
3	Welk product staat op de binnendoos/plasticzak? Medical Face Mask
4	Welke leverancier staat op de binnendoos/plasticzak? Yongli
5	Staat een CE merk op de doos/plasticzak? ☐ Ja ☒ Nee
6	Welke norm(en) staan op de binnendoos/plasticzak? ☐ EN 149 (FFP) ☐ N95 ☐ KN95 ☐ P2 (Australia) ☐ KMOEL- 2017-64 Korea ☐ DS (Japan JMHLW-- ☒ EN 14683 (Surgical) ☐ YY 0469-2011 (Surgical) ☐ YY/T 0969 (Face mask) ☐ GB2626-2006 ☐ Anders:.... ☐ Geen
7	Is er een instructie bijgeleverd? ☒ Staat op de doos ☐ Los in de doos ☐ Ontbreekt
8	Heeft de instructie betrekking op het masker in de doos/plasticzak? ☒ Ja ☐ Nee ☒ N.v.t.
9	Staat er een expiratie datum op binnendoos/plasticzak (shelf life max. 5 jaar) ☒ Ja ☐ Nee 2022-05

Bijzonderheden:

VERPAKKING OM HET INDIVUELE MASKER	
1	Zit er verpakking om het masker? ☐ Ja ☒ Nee
	Zo ja:
2	Hoe is de verpakking te openen? ☐ Gemakkelijk ☐ Lastig ☐ Moeilijk
3	Correspondeert de verpakking met de informatie op de binnendoos? ☐ Ja ☐ Nee, het volgende wijkt af:

Bijzonderheden:

HET MASKER	
1	Algemene indruk masker? ☒ Integriteit in orde ☐ Integriteit niet in orde, omdat:
2	Komt het masker overeen met de foto/tekening op de binnendoos? ☒ Ja ☐ Ontbreekt ☐ Nee, het volgende wijkt af:
3	Wat staat er op het masker: <i>hvt</i>
	Komt dit overeen met wat op de verpakking van het masker staat? ☐ Ja ☐ Nee ☒ N.v.t.
	Komt dit overeen met wat er op de binnendoos staat? ☐ Ja ☐ Nee ☒ N.v.t.
4	Is de stof is non woven? ☒ Ja ☐ Nee
5	Aantal lagen van het masker: ☐ 1 ☐ 2 ☒ 3 ☐ 4 ☐ 5
6	Heeft het masker een nare geur? ☐ Ja ☐ Nee
7	Hoe wordt masker op hoofd bevestigd? ☐ Met hoofdbanden ☒ Met oor elastieken ☐ Losse touwtjes
8	Hoe zitten de touwtjes/elastieken aan het masker vast? ☒ Erop gestanst ☐ Erin gestanst ☐ Gelijmd ☐ Genaaid ☐ Vast d.m.v.
9	Wat is de kwaliteit van de bevestiging van de touwtjes/elastieken? ☒ Goed stevig vast ☐ Bij een keer dragen geen probleem ☐ Bij extra kracht laat het los
10	Neusstuk ☒ Metaal in de stof ☐ Kunststof met metalen kern in de stof ☐ Kunststof in de stof (<i>eerder in andere hits aangelijfd</i>) ☐ Metaal aan de buitenkant ☐ Met filterdoek afgedekt ☐ Rubber strip ☐

Bijzonderheden:

DE FIT VAN HET MASKER OP HET GEZICHT		
1	Hoe past het masker op het gezicht?	☐ Goed ☐ Ruimte bij de kin ☒ Ruimte bij neusstuk na aandrukken ☐ Slechte aansluiting bij wangen ☐
2	Waar ontsnapt lucht bij de leaktest	☐ Niet ☒ Neus ☐ Kin ☒ Zijkant ☐ N.v.t. (Chirurgisch)
3	Kan het masker met de touwtjes en elastieken strakker op het gezicht worden getrokken?	☒ Ja ☒ Nee ☐ N.v.t.
4	Wat gebeurt er als het gezicht beweegt?	☐ Masker blijft op zijn plek ☒ Masker verschuift ☐
5	Behoudt de neusklem zijn vorm gedurende circa 1 min?	☒ Ja ☐ Nee ☐ N.v.t.
6	Chirurgisch masker: Wat is de hoogte van het masker wanneer deze uitgevouwen is?	15,5
7	Chirurgisch masker: wat is de lengte van de neusklem?	10,5

Bij een chirurgisch masker			
Doos	Materiaal neusbeugel	Breedte neusbeugel	Lengte van het uitgeklapte masker
1			
2			
3	1970-3	met hand in plastic	15,5
4			
5			
6			
7			
8			
9			
10			
11			
12			

Bijzonderheden:

Disclaimer: Deze rapportage bevat een onafhankelijk indicatief oordeel van de kwaliteit van de producten. Dit betekent dat deze resultaten slechts een gedeeltelijke weergave zijn van de kwaliteit van de producten. De rapportage is alleen te gebruiken door het LCH en het ministerie VWS ter ondersteuning bij het vrijgeven van producten op de markt. Gegevens uit deze rapportage mogen niet gedeeld worden met derden. Indien u niet de geadresseerde bent of dit bericht abusievelijk aan u is verzonden, wordt u verzocht dat aan de afzender te melden en de rapportage te verwijderen. Het LCH en het ministerie VWS aanvaarden geen aansprakelijkheid voor schade, van welke aard ook, die verband houdt met risico's verbonden aan het elektronisch verzenden van berichten.

Beoordelingsformulier voor de spattest

Datum	29-06-2020
Inkoopnummer (indien bekend)	
Lotnummer/X-nummer/Z-nummer	1770-6
Leverancier/Afzender	
Productnaam of naam fabrikant	
Type mondmasker (volgens de fabrikant)	

Conditionering (indien van toepassing):

Conditionering uitgevoerd	☒ Ja	☐ Nee
Aanvang conditionering (Datum & tijd)	29-06 8:15	
Einde conditionering (Datum & tijd)	29-06 13:49	

Spattest

Er worden bij voorkeur 32, maar minimaal 16, maskers getest. Maximaal 3 (32 maskers), en 2 (16 maskers) fails worden geaccepteerd voor een pass.

Masker	Pass	Fail	Masker	Pass	Fail
1	☒	☐	17	☒	☐
2	☒	☐	18	☒	☐
3	☒	☐	19	☒	☐
4	☒	☐	20	☒	☐
5	☒	☐	21	☒	☐
6	☒	☐	22	☒	☐
7	☒	☐	23	☒	☐
8	☒	☐	24	☒	☐
9	☒	☐	25	☒	☐
10	☒	☐	26	☒	☐
11	☒	☐	27	☒	☐
12	☒	☐	28	☒	☐
13	☒	☐	29	☒	☐
14	☒	☐	30	☒	☐
15	☒	☐	31	☒	☐
16	☒	☐	32	☒	☐


Aantal gemeten maskers: 32

Aantal fails: 0

Uitslag Spattest

☒ PASS

☐ FAIL



Beoordelingsformulier voor de spattest

Datum	29-06-2020
Inkoopnummer (indien bekend)	
Lotnummer/X-nummer/Z-nummer	1770-7
Leverancier/Afzender	
Productnaam of naam fabrikant	
Type mondmasker (volgens de fabrikant)	

Conditionering (indien van toepassing):

Conditionering uitgevoerd	☒ Ja	☐ Nee
Aanvang conditionering (Datum & tijd)	29-06 8:15	
Einde conditionering (Datum & tijd)	29-06 14:10	

Spattest

Er worden bij voorkeur 32, maar minimaal 16, maskers getest. Maximaal 3 (32 maskers), en 2 (16 maskers) fails worden geaccepteerd voor een pass.

Masker	Pass	Fail	Masker	Pass	Fail
1	☒	☐	17	☒	☐
2	☒	☐	18	☒	☐
3	☒	☐	19	☒	☐
4	☐	☒	20	☒	☐
5	☒	☐	21	☒	☐
6	☒	☐	22	☒	☐
7	☒	☐	23	☒	☐
8	☒	☐	24	☒	☐
9	☒	☐	25	☒	☐
10	☒	☐	26	☒	☐
11	☒	☐	27	☒	☐
12	☒	☐	28	☒	☐
13	☒	☐	29	☒	☐
14	☒	☐	30	☒	☐
15	☒	☐	31	☒	☐
16	☒	☐	32	☒	☐

Aantal gemeten maskers: 32

Aantal fails: 0

Uitslag Spattest

☒ PASS

☐ FAIL



Beeoordelingsformulier voor de spattest

Datum	25-6-2020
Inkoopnummer (indien bekend)	
Lotnummer/X-nummer/Z-nummer	1770-1
Leverancier/Afzender	
Productnaam of naam fabrikant	
Type mondkmasker (volgens de fabrikant)	

Conditionering (indien van toepassing):

Conditionering uitgevoerd	☒ Ja	☐ Nee
Aanvang conditionering (Datum & tijd)	24-6	
Einde conditionering (Datum & tijd)	25-6 15:45	

Spattest

Er worden bij voorkeur 32, maar minimaal 16, maskers getest. Maximaal 3 (32 maskers), en 2 (16 maskers) fails worden geaccepteerd voor een pass.

Masker	Pass	Fail	Masker	Pass	Fail
1	☒	☐	17	☒	☐
2	☒	☐	18	☒	☐
3	☒	☐	19	☒	☐
4	☒	☐	20	☒	☐
5	☒	☐	21	☒	☐
6	☒	☐	22	☒	☐
7	☒	☐	23	☒	☐
8	☒	☐	24	☒	☐
9	☒	☐	25	☒	☐
10	☒	☐	26	☒	☐
11	☒	☐	27	☒	☐
12	☒	☐	28	☒	☐
13	☒	☐	29	☒	☐
14	☒	☐	30	☒	☐
15	☒	☐	31	☒	☐
16	☒	☐	32	☒	☐

Aantal gemeten maskers: 32

Aantal fails: 0

Uitslag Spattest

☒ PASS

☐ FAIL

1. The first part of the document is a list of names and addresses of the members of the committee.

2. The second part of the document is a list of names and addresses of the members of the committee.

3. The third part of the document is a list of names and addresses of the members of the committee.

4. The fourth part of the document is a list of names and addresses of the members of the committee.

5. The fifth part of the document is a list of names and addresses of the members of the committee.

6. The sixth part of the document is a list of names and addresses of the members of the committee.

7. The seventh part of the document is a list of names and addresses of the members of the committee.

8. The eighth part of the document is a list of names and addresses of the members of the committee.

9. The ninth part of the document is a list of names and addresses of the members of the committee.

10. The tenth part of the document is a list of names and addresses of the members of the committee.

11. The eleventh part of the document is a list of names and addresses of the members of the committee.

12. The twelfth part of the document is a list of names and addresses of the members of the committee.

13. The thirteenth part of the document is a list of names and addresses of the members of the committee.

14. The fourteenth part of the document is a list of names and addresses of the members of the committee.

15. The fifteenth part of the document is a list of names and addresses of the members of the committee.

16. The sixteenth part of the document is a list of names and addresses of the members of the committee.

17. The seventeenth part of the document is a list of names and addresses of the members of the committee.

18. The eighteenth part of the document is a list of names and addresses of the members of the committee.

19. The nineteenth part of the document is a list of names and addresses of the members of the committee.

20. The twentieth part of the document is a list of names and addresses of the members of the committee.

21. The twenty-first part of the document is a list of names and addresses of the members of the committee.

22. The twenty-second part of the document is a list of names and addresses of the members of the committee.

23. The twenty-third part of the document is a list of names and addresses of the members of the committee.

24. The twenty-fourth part of the document is a list of names and addresses of the members of the committee.

25. The twenty-fifth part of the document is a list of names and addresses of the members of the committee.

26. The twenty-sixth part of the document is a list of names and addresses of the members of the committee.

27. The twenty-seventh part of the document is a list of names and addresses of the members of the committee.

28. The twenty-eighth part of the document is a list of names and addresses of the members of the committee.

29. The twenty-ninth part of the document is a list of names and addresses of the members of the committee.

Beoordelingsformulier voor de spattest

Datum	29-06-2020
Inkoopnummer (indien bekend)	
Lotnummer/X-nummer/Z-nummer	1770-3
Leverancier/Afzender	
Productnaam of naam fabrikant	
Type mondmasker (volgens de fabrikant)	

Conditionering (indien van toepassing):

Conditionering uitgevoerd	☒ Ja	☐ Nee
Aanvang conditionering (Datum & tijd)	29-06 8:15	
Einde conditionering (Datum & tijd)	29-06 12:40	

Spattest

Er worden bij voorkeur 32, maar minimaal 16, maskers getest. Maximaal 3 (32 maskers), en 2 (16 maskers) fails worden geaccepteerd voor een pass.

Masker	Pass	Fail	Masker	Pass	Fail
1	☒	☐	17	☒	☐
2	☒	☐	18	☒	☐
3	☒	☐	19	☒	☐
4	☒	☐	20	☒	☐
5	☒	☐	21	☒	☐
6	☒	☐	22	☒	☐
7	☒	☐	23	☒	☐
8	☒	☐	24	☒	☐
9	☒	☐	25	☒	☐
10	☒	☐	26	☒	☐
11	☒	☐	27	☒	☐
12	☒	☐	28	☒	☐
13	☒	☐	29	☒	☐
14	☒	☐	30	☒	☐
15	☒	☐	31	☒	☐
16	☒	☐	32	☒	☐

Aantal gemeten maskers: 32

Aantal fails: 0

Uitslag Spattest

☒ PASS

☐ FAIL



Beoordelingsformulier voor de spattest

Datum	29-06-2020
Inkoopnummer (indien bekend)	
Lotnummer/X-nummer/Z-nummer	1770-4
Leverancier/Afzender	
Productnaam of naam fabrikant	
Type mondmasker (volgens de fabrikant)	

Conditionering (indien van toepassing):

Conditionering uitgevoerd	☒ Ja	☐ Nee
Aanvang conditionering (Datum & tijd)	29-06 8:15	
Einde conditionering (Datum & tijd)	13:20	

Spattest

Er worden bij voorkeur 32, maar minimaal 16, maskers getest. Maximaal 3 (32 maskers), en 2 (16 maskers) fails worden geaccepteerd voor een pass.

Masker	Pass	Fail	Masker	Pass	Fail
1	☒	☐	17	☒	☐
2	☒	☐	18	☒	☐
3	☒	☐	19	☒	☐
4	☒	☐	20	☒	☐
5	☒	☐	21	☒	☐
6	☒	☐	22	☒	☐
7	☒	☐	23	☒	☐
8	☒	☐	24	☒	☐
9	☒	☐	25	☒	☐
10	☒	☐	26	☒	☐
11	☒	☐	27	☒	☐
12	☒	☐	28	☒	☐
13	☒	☐	29	☒	☐
14	☒	☐	30	☒	☐
15	☒	☐	31	☒	☐
16	☒	☐	32	☒	☐


Aantal gemeten maskers: 32

Aantal fails: 0

Uitslag Spattest

☒ PASS

☐ FAIL



Beoordelingsformulier voor de spattest

Datum	25-6-2020
Inkoopnummer (indien bekend)	
Lotnummer/X-nummer/Z-nummer	1770-5
Leverancier/Afzender	
Productnaam of naam fabrikant	
Type mondmasker (volgens de fabrikant)	

Conditionering (indien van toepassing):

Conditionering uitgevoerd	☒ Ja	☐ Nee
Aanvang conditionering (Datum & tijd)	24-6	
Einde conditionering (Datum & tijd)	25-6 15:20	

Spattest

Er worden bij voorkeur 32, maar minimaal 16, maskers getest. Maximaal 3 (32 maskers), en 2 (16 maskers) fails worden geaccepteerd voor een pass.

Masker	Pass	Fail	Masker	Pass	Fail
1	☒	☐	17	☒	☐
2	☒	☐	18	☒	☐
3	☒	☐	19	☒	☐
4	☒	☐	20	☒	☐
5	☒	☐	21	☒	☐
6	☒	☐	22	☒	☐
7	☒	☐	23	☒	☐
8	☒	☐	24	☒	☐
9	☒	☐	25	☒	☐
10	☒	☐	26	☒	☐
11	☒	☐	27	☒	☐
12	☒	☐	28	☒	☐
13	☒	☐	29	☒	☐
14	☒	☐	30	☒	☐
15	☒	☐	31	☒	☐
16	☒	☐	32	☒	☐

Aantal gemeten maskers: 32

Aantal fails: 0

Uitslag Spattest

☒ PASS

☐ FAIL



Inboekrapport VWS

Naam Ontvanger:

5.1.2e

Datum:

16-06-20

Aantal

inkooporder/pakbon

Lotnummer

1770

Aantal pallets

25

Totaal stuks

1.000.000

Aantal samples

450

Totaal aantal na samples

999.550

Aantal per pallet

40.000

Aantal per doos

2000

Aantal per verpakkingseenheid

50

Aantal komt overeen
met
inkooporder/pakbonJA /
NEE

Zo niet

Telling bevestigd door

Naam:

Paraaf:

Artikelnummer:

g 74

Inkooporder
nummer

304374

LCH-nummer

2020-0200-07122

Kwaliteitscontrole
RIVMJA /
NEE

GOEDGEKEURD / AFGEKEURD

Bijzonderheden:

Locatie:

013-149-05 1m

013-197-05

ingeboekt
door:



nutrilab

Certificate of analysis

Order-no : 840568

RIVM - (Corona Advies)

A.van Leeuwenhoeklaan 9
3721 MA BILTHOVEN

Description	: 1770.1	Sample date	:
Client project no	: 20200626	Start date microbiology	: 29 06 2020
Sample received	: 25 06 2020	Sample delivery	: Nutrilab BV
Report date	: 30 06 2020	Sample temperature	: Room temperature
Sampling	:	Sample condition	: Sample and packing intact
Packing	:		
Sealed	: No		

Test	Result
Bacterial Filtration Efficiency (BFE)	
I 35000 Testing BFE (n=5) (equiv. NEN EN 14683+C1)	
I 35050 Test conditions	
I 35051 Dimensions of test specimens (width x height)	Whole mask
I 35052 Size of the area tested (width x height)	48.0 cm ²
I 35053 Side facing the aerosol	Face side
I 35054 Flow rate during testing	28.3 L/min
I 35060 Mean of the total plate counts of the two positive controls	2.221.5 cfu
I 35070 Mean plate count of the negative controls	0 cfu
I 35100 Bacterial Filtration Efficiency (BFE, equiv. NEN EN 14683+C1)	
I 35101 BFE specimen 1	99.9 %
I 35102 BFE specimen 2	100.0 %
I 35103 BFE specimen 3	100.0 %
I 35104 BFE specimen 4	100.0 %
I 35105 BFE specimen 5	100.0 %
I 35199 Average BFE	100.0 %
Conclusion*	
I 35290 Mask type based on BFE performance requirements for medical face masks	Ia / II / IIR







nutrilab

Certificate of analysis

Order-no : 840568

*) Performance requirements for medical face masks (acc. European Standard no. EN 14683:2019+AC):

Bacterial filtration efficiency (BFE) (%): Type I a ≥ 95 , Type II ≥ 98 , Type IIR ≥ 98
 Differential pressure (Pa/cm²): Type I a < 40 , Type II < 40 , Type IIR < 60
 Splash resistance pressure (kPa): Type I a Not required, Type II Not required, Type IIR ≥ 16.0
 Microbial cleanliness (cfu/g): Type I a ≤ 30 , Type II ≤ 30 , Type IIR ≤ 30

a) Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

This certificate of analysis is a test report. The tested samples are part of the mentioned batch/lot number. Batch validation is not the scope of this report.

Start date analysis: 29-06-2020, end date: 30-06-2020.

Findings are based on the sample as submitted. For more detailed information on applied methods please contact the operational manager.

Any interpretation of analytical results mentioned on this certificate lies outside the scope of accreditation.

With the unit % is meant, w/w%, unless otherwise stated.

Nutrilab is not responsible for the information provided by the client.

Only reproduction of the entire certificate is permitted without explicit written authorisation by Nutrilab BV.

5.1.2e

5.1.2e

Nutrilab bv
 Postbus 7, 4284 ZG Rijswijk
 Burgstraat 12, 4283 GG Giessen

020 76000000 info@nutrilab.nl

www.nutrilab.nl

KvK 18114291


BTW 5.1.5



Page 2 / 2

Explanation of symbols

Q RvA accredited test (ISO / IEC 17025)
 I Test performed by Nutrilab BV
 E Test performed by sub contractor





nutrilab

Certificate of analysis

Order no : 841305

RIVM - (Corona Advies)

A. van Leeuwenhoeklaan 9
3721 MA BILTHOVEN

Description	:	1770 3	Sample date	:	
Client project no	:	20200630	Start date microbiology	:	30 06 2020
Sample received	:	30 06 2020	Sample delivery	:	
Report date	:	01 07 2020	Sample temperature	:	
Sampling	:		Sample condition	:	Sample and packing intact
Packing	:				
Sealed	:	No			

Test	Result
Bacterial Filtration Efficiency (BFE)	
I 35000 Testing BFE (n=5) (equiv. NEN EN 14683+C1)	
I 35050 Test conditions	
I 35051 Dimensions of test specimens (width x height)	Whole mask
I 35052 Size of the area tested (width x height)	48.0 cm ²
I 35053 Side facing the aerosol	Face side
I 35054 Flow rate during testing	28.3 L/min
I 35060 Mean of the total plate counts of the two positive controls	2.693 cfu
I 35070 Mean plate count of the negative controls	0 cfu
I 35100 Bacterial Filtration Efficiency (BFE, equiv. NEN EN 14683+C1)	
I 35101 BFE specimen 1	100 %
I 35102 BFE specimen 2	100 %
I 35103 BFE specimen 3	100 %
I 35104 BFE specimen 4	100 %
I 35105 BFE specimen 5	100 %
I 35199 Average BFE	100 %
Conclusion*	
I 35290 Mask type based on BFE performance requirements for medical face masks	Ia / II / IIR

Nutrilab bv
Postbus 7, 4284 ZG Rijswijk
Burgstraat 12, 4283 GG Giessen

5.1.2e

KvK 18114291

BTW 5.1.5



Page 1 / 2

Explanation of symbols:

Q RvA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub contractor



nutrilab

Certificate of analysis

Order-no : 841305

*) Performance requirements for medical face masks (acc. European Standard no. EN 14683:2019+AC):

Bacterial filtration efficiency (BFE) (%): Type I a ≥ 95 , Type II ≥ 98 , Type IIR ≥ 98

Differential pressure (Pa/cm²): Type I a < 40 , Type II < 40 , Type IIR < 60

Splash resistance pressure (kPa): Type I a Not required, Type II Not required, Type IIR ≥ 16.0

Microbial cleanliness (cfu/g): Type I a ≤ 30 , Type II ≤ 30 , Type IIR ≤ 30

a) Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

This certificate of analysis is a test report. The tested samples are part of the mentioned batch/lot number. Batch validation is not the scope of this report.

Start date analysis: 30 06 2020, end date: 01 07 2020.

Findings are based on the sample as submitted. For more detailed information on applied methods please contact the operational manager.

Any interpretation of analytical results mentioned on this certificate lies outside the scope of accreditation.

With the unit % is meant w/w%, unless otherwise stated.

Nutrilab is not responsible for the information provided by the client.

Only reproduction of the entire certificate is permitted without explicit written authorisation by Nutrilab BV.

5.1.2e

5.1.2e

Nutrilab bv

Postbus 7, 4284 ZG, Rijswijk

Burgstraat 12, 4283 GG, Giessen

5.1.2e

onele.em@nutrilab.nl

www.nutrilab.nl

KvK 18114291

BTW 5.1.5



Page 2 / 2

Explanation of symbols:

Q RVA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub contractor



nutrilab

Certificate of analysis

Order-no : 841307

RIVM - (Corona Advies)

A.van Leeuwenhoeklaan 9
3721 MA BILTHOVEN

Description	: 1770 4	Sample date	:	
Client project no	: 20200630	Start date microbiology	:	01 07 2020
Sample received	: 30 06 2020	Sample delivery	:	
Report date	: 02 07 2020	Sample temperature	:	
Sampling	:	Sample condition	:	Sample and packing intact
Packing	:			
Sealed	: No			

Test	Result
Bacterial Filtration Efficiency (BFE)	
I 35000 Testing BFE (n=5) (equiv. NEN EN 14683+C1)	
I 35050 Test conditions	
I 35051 Dimensions of test specimens (width x height)	Whole mask
I 35052 Size of the area tested (width x height)	48.0 cm ²
I 35053 Side facing the aerosol	Face side
I 35054 Flow rate during testing	28.3 L/min
I 35060 Mean of the total plate counts of the two positive controls	2,977 cfu
I 35070 Mean plate count of the negative controls	0 cfu
I 35100 Bacterial Filtration Efficiency (BFE, equiv. NEN EN 14683+C1)	
I 35101 BFE specimen 1	100 %
I 35102 BFE specimen 2	100 %
I 35103 BFE specimen 3	100 %
I 35104 BFE specimen 4	100 %
I 35105 BFE specimen 5	100 %
I 35199 Average BFE	100 %
Conclusion*	
I 35290 Mask type based on BFE performance requirements for medical face masks	I/II/III

Nutrilab bv
Postbus 7, 4284 ZG Rijswijk
Burgstraat 12, 4283 CG Giessen

5.1.2014
www.nutrilab.nl

KvK 18114291
BTW 5.1.5



Page 1 / 2

Explanation of symbols

Q RvA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub contractor

5.1.2e



nutrilab

Certificate of analysis

Order no: 841307

*) Performance requirements for medical face masks (acc. European Standard no. EN 14683:2019+AC):

Bacterial filtration efficiency (BFE) (%): Type I a ≥ 95 , Type II ≥ 98 , Type IIR ≥ 98

Differential pressure (Pa/cm²): Type I a < 40 , Type II < 40 , Type IIR < 60

Splash resistance pressure (kPa): Type I a Not required, Type II Not required, Type IIR ≥ 16.0

Microbial cleanliness (cfu/g): Type I a ≤ 30 , Type II ≤ 30 , Type IIR ≤ 30

a) Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements

This certificate of analysis is a test report. The tested samples are part of the mentioned batch/lot number. Batch validation is not the scope of this report.

Start date analysis: 01-07-2020, end date: 02-07-2020.

Findings are based on the sample as submitted. For more detailed information on applied methods please contact the operational manager.

Any interpretation of analytical results mentioned on this certificate lies outside the scope of accreditation.

With the unit % is meant w/w%, unless otherwise stated.

Nutrilab is not responsible for the information provided by the client.

Only reproduction of the entire certificate is permitted without explicit written authorisation by Nutrilab BV.

5.1.2e

5.1.2e

Nutrilab bv
Postbus 7, 4284 ZG Rijswijk
Burgstraat 12, 4283 GG Giessen

info@nutrilab.nl

www.nutrilab.nl

KvK 18114291

BTW 5.1.5



Page 2 / 2

Explanation of symbols

Q RvA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
F Test performed by sub contractor

5.1.2e



nutrilab

Certificate of analysis

Order-no : 840569

RIVM - (Corona Advies)

A.van Leeuwenhoeklaan 9
3721 MA BILTHOVEN

Description	: 1770 5	Sample date	: 30 06 2020
Client project no	: 20200626	Start date microbiology	: 30 06 2020
Sample received	: 25 06 2020	Sample delivery	: Nutrilab BV
Report date	: 01 07 2020	Sample temperature	: Room temperature
Sampling	:	Sample condition	: Sample and packing intact
Packing	:		
Sealed	: No		

Test	Result
Bacterial Filtration Efficiency (BFE)	
I 35000 Testing BFE (n=5) (equiv. NEN-EN 14683+C1)	
I 35050 Test conditions	
I 35051 Dimensions of test specimens (width x height)	Whole mask
I 35052 Size of the area tested (width x height)	48.0 cm ²
I 35053 Side facing the aerosol	Face side
I 35054 Flow rate during testing	28.3 L/min
I 35060 Mean of the total plate counts of the two positive controls	2,693 cfu
I 35070 Mean plate count of the negative controls	0 cfu
I 35100 Bacterial Filtration Efficiency (BFE, equiv. NEN-EN 14683+C1)	
I 35101 BFE specimen 1	100 %
I 35102 BFE specimen 2	100 %
I 35103 BFE specimen 3	100 %
I 35104 BFE specimen 4	100 %
I 35105 BFE specimen 5	100 %
I 35199 Average BFE	100 %
Conclusion*	
I 35290 Mask type based on BFE performance requirements for medical face masks	Ia / II / IIR

Nutrilab bv
Postbus 7, 4284 ZG Rijswijk
Burgstraat 12, 4283 GG Giessen

5.1.2e
5.1.5



Page 1 / 2

Explanation of symbols:

Q RvA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub-contractor



nutrilab

Certificate of analysis

Order-no : 840569

*) Performance requirements for medical face masks (acc. European Standard no. EN 14683:2019+AC):

Bacterial filtration efficiency (BFE) (%): Type I a ≥ 95 , Type II ≥ 98 , Type IIR ≥ 98

Differential pressure (Pa/cm²): Type I a < 40 , Type II < 40 , Type IIR < 60

Splash resistance pressure (kPa): Type I a Not required, Type II Not required, Type IIR ≥ 16.0

Microbial cleanliness (cfu/g): Type I a ≤ 30 , Type II ≤ 30 , Type IIR ≤ 30

a) Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

This certificate of analysis is a test report. The tested samples are part of the mentioned batch/lot number. Batch validation is not the scope of this report.

Start date analysis: 30 06 2020, end date: 01 07 2020.

Findings are based on the sample as submitted. For more detailed information on applied methods please contact the operational manager.

Any interpretation of analytical results mentioned on this certificate lies outside the scope of accreditation.

With the unit % is meant, w/w%, unless otherwise stated.

Nutrilab is not responsible for the information provided by the client.

Only reproduction of the entire certificate is permitted without explicit written authorisation by Nutrilab BV.

5.1.2e

5.1.2e

Nutrilab bv
Postbus 7, 4284 ZG Rijnswijk
Burgstraat 12, 4283 GG Giessen

5.1.2e

phone: +31 (0) 475 411111
www.nutrilab.nl

KvK 18114291

BTW 5.1.5



Page 2 / 2

Explanation of symbols:

Q RVA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub-contractor



nutrilab

Certificate of analysis

Order-no : 841308

RIVM - (Corona Advies)

A.van Leeuwenhoeklaan 9
3721 MA BILTHOVEN

Description	:	1770 6	Sample date	:	
Client project no	:	20200630	Start date microbiology	:	01-07-2020
Sample received	:	30-06-2020	Sample delivery	:	
Report date	:	02-07-2020	Sample temperature	:	
Sampling	:		Sample condition	:	Sample and packing intact
Packing	:				
Sealed	:	No			

Test	Result
Bacterial Filtration Efficiency (BFE)	
I 35000 Testing BFE (n=5) (equiv. NEN EN 14683+C1)	
I 35050 Test conditions	
I 35051 Dimensions of test specimens (width x height)	Whole mask
I 35052 Size of the area tested (width x height)	48.0 cm ²
I 35053 Side facing the aerosol	Face side
I 35054 Flow rate during testing	28.3 L/min
I 35060 Mean of the total plate counts of the two positive controls	2.977 cfu
I 35070 Mean plate count of the negative controls	0 cfu
I 35100 Bacterial Filtration Efficiency (BFE, equiv. NEN EN 14683+C1)	
I 35101 BFE specimen 1	100 %
I 35102 BFE specimen 2	100 %
I 35103 BFE specimen 3	100 %
I 35104 BFE specimen 4	100 %
I 35105 BFE specimen 5	100 %
I 35199 Average BFE	100 %
Conclusion*	
I 35290 Mask type based on BFE performance requirements for medical face masks	I/II/III

Nutrilab bv
Postbus 7, 4284 ZG Rijswijk
Burgstraat 12, 4283 GG Giessen

t. 5.1.2e
f. 5.1.5

KvK 18114291
BTW 5.1.5



Page 1 / 2

Explanation of symbols

Q RvA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub contractor



nutrilab

Certificate of analysis

Order no: 841308

*) Performance requirements for medical face masks (acc. European Standard no. EN 14683:2019+AC):

Bacterial filtration efficiency (BFE) (%): Type I a ≥ 95 , Type II ≥ 98 , Type IIR ≥ 98

Differential pressure (Pa/cm²): Type I a < 40 , Type II < 40 , Type IIR < 60

Splash resistance pressure (kPa): Type I a Not required, Type II Not required, Type IIR ≥ 16.0

Microbial cleanliness (cfu/g): Type I a ≤ 30 , Type II ≤ 30 , Type IIR ≤ 30

a) Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

This certificate of analysis is a test report. The tested samples are part of the mentioned batch/lot number. Batch validation is not the scope of this report.

Start date analysis: 01-07-2020, end date: 02-07-2020

Findings are based on the sample as submitted. For more detailed information on applied methods please contact the operational manager.

Any interpretation of analytical results mentioned on this certificate lies outside the scope of accreditation.

With the unit % is meant, w/w%, unless otherwise stated.

Nutrilab is not responsible for the information provided by the client.

Only reproduction of the entire certificate is permitted without explicit written authorisation by Nutrilab BV.

5.1.2e

5.1.2e

Nutrilab bv
Postbus 7, 4284 ZG Rijswijk
Burgstraat 12, 4283 GG Giessen

5.1.2e

info@nutrilab.nl
www.nutrilab.nl

KvK 18114291

BTW 5.1.5



Page 2 / 2

Explanation of symbols

Q RVA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub-contractor



nutrilab

Certificate of analysis

Order-no : 841309

RIVM - (Corona Advies)

A. van Leeuwenhoeklaan 9
3721 MA BILTHOVEN

Description	: 1770.7	Sample date	:	
Client project no	: 20200630	Start date microbiology	:	01-07-2020
Sample received	: 30.06.2020	Sample delivery	:	
Report date	: 02.07.2020	Sample temperature	:	
Sampling	:	Sample condition	:	Sample and packing intact
Packing	:			
Sealed	: No			

Test	Result
Bacterial Filtration Efficiency (BFE)	
I 35000 Testing BFE (n=5) (equiv. NEN EN 14683+C1)	
I 35050 Test conditions	
I 35051 Dimensions of test specimens (width x height)	Whole mask
I 35052 Size of the area tested (width x height)	48.0 cm ²
I 35053 Side facing the aerosol	Face side
I 35054 Flow rate during testing	28.3 L/min
I 35060 Mean of the total plate counts of the two positive controls	2,977 cfu
I 35070 Mean plate count of the negative controls	0 cfu
I 35100 Bacterial Filtration Efficiency (BFE, equiv. NEN EN 14683+C1)	
I 35101 BFE specimen 1	100 %
I 35102 BFE specimen 2	100 %
I 35103 BFE specimen 3	100 %
I 35104 BFE specimen 4	100 %
I 35105 BFE specimen 5	100 %
I 35199 Average BFE	100 %
Conclusion*	
I 35290 Mask type based on BFE performance requirements for medical face masks	I/II/IIR

Nutrilab bv
Postbus 7, 4284 ZG Rijswijk
Burgstraat 12, 4283 GG Giessen

5.1.2e

nutrilab.nl
www.nutrilab.nl

KvK 18114291
BTW 5.1.5



Page 1 / 2

Explanation of symbols

Q RVA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub contractor



nutrilab

Certificate of analysis

Order-no : 841309

*) Performance requirements for medical face masks (acc. European Standard no. EN 14683:2019+AC):

Bacterial filtration efficiency (BFE) (%): Type I a ≥ 95 , Type II ≥ 98 , Type IIR ≥ 98

Differential pressure (Pa/cm²): Type I a < 40 , Type II < 40 , Type IIR < 60

Splash resistance pressure (kPa): Type I a Not required, Type II Not required, Type IIR ≥ 16.0

Microbial cleanliness (cfu/g): Type I a ≤ 30 , Type II ≤ 30 , Type IIR ≤ 30

a) Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

This certificate of analysis is a test report. The tested samples are part of the mentioned batch/lot number. Batch validation is not the scope of this report.

Start date analysis: 01-07-2020, end date: 02-07-2020.

Findings are based on the sample as submitted. For more detailed information on applied methods please contact the operational manager.

Any interpretation of analytical results mentioned on this certificate lies outside the scope of accreditation.

With the unit % is meant: w/w%, unless otherwise stated.

Nutrilab is not responsible for the information provided by the client.

Only reproduction of the entire certificate is permitted without explicit written authorisation by Nutrilab BV.

5.1.2e

5.1.2e

Nutrilab bv

Postbus 7, 4284 ZG Rijswijk

Burgstraat 12, 4283 GG Giessen

t.

f.

5.1.2e

phone: +31 (0) 78 611 1429

www.nutrilab.nl

KvK 18114291

BTW 5.1.5



Page 2 / 2

Explanation of symbols

Q RvA accredited test (ISO / IEC 17025)

I Test performed by Nutrilab BV

E Test performed by sub-contractor

Inboekrapport VWS

Naam Ontvanger:

5.1.2e

Datum:

16-06-20

Aantal
inkooporder/pakbon

Lotnummer

1770

Aantal pallets

25

Totaal stuks

1.000.000

Aantal samples

450

Totaal aantal na samples

999.550

Aantal per pallet

40.000

Aantal per doos

2000

Aantal per verpakkingseenheid

50

Aantal komt overeen
met
inkooporder/pakbonJA /
NEE

Zo niet

Telling bevestigd door

Naam:

Paraaf:

Artikelnummer:

g 74

Inkooporder
nummer

304374

LCH-nummer

2020-0200-07202

Kwaliteitscontrole
RIVMJA /
NEE

GOEDGEKEURD / AFGEKEURD

Bijzonderheden:

Locatie:

013-149-05 +m
013-197-05ingeboekt
door:

1-15 y compris et / inclusief
 19+20+21 Las partes encadradas de las partes grises deben ser rellenas por el transportista / De dik omlijnde vakken moeten ingevuld worden
 19+20+21 worden door de vervoerder / Die mit fett gedruckten Linien umgrenzten Rubriken müssen vom Frachtführer ausgefüllt werden

1-15 y compris et / inclusief
 19+20+21 worden door de vervoerder / Die mit fett gedruckten Linien umgrenzten Rubriken müssen vom Frachtführer ausgefüllt werden

A remplir sous la responsabilité de l'expéditeur / In te vullen onder verantwoordelijkheid van de afzender / Ausfüllen unter der Verantwortung des Absenders

2 Exemplaire pour destinataire / 2 Exemplare für Empfänger		LETTRE DE VOITURE - DOCUMENT DE VRACHTBRIEF - VERVOERDDOCUMENT FRACHTBRIEF - TRANSPORTDOKUMENT		CMR Code transporteur / Vervoerderscode No / Code Frachtführer		
1 Expéditeur (nom, adresse, pays) / Afzender (naam, adres, land) Absender (Name, Anschrift, Land)		Ce transport est soumis, nonobstant toute clause contraire, à la Convention relative au contrat de transport international de marchandises par route (CMR). Dit vervoer is, ongeacht enig tegengesteld beding, onderworpen aan het CMR-Verdrag. Diese Beförderung unterliegt trotz einer gegenseitigen Abmachung den Bestimmungen des Übereinkommens über den Beförderungsvertrag im internationalen Straßengüterverkehr (CMR).				
NTG Air & Ocean Netherlands BV Shannonweg 13 1118 LA Schiphol The Netherlands		3120.20.2.00229				
2 Destinataire (nom, adresse, pays) / Geadresseerde (naam, adres, land) Empfänger (Name, Anschrift, Land)		16 Transporteur (nom, adresse, pays) / Vervoederer (naam, adres, land) Frachtführer (Name, Adresse, Land)				
Mediq Nederland BV 5.1.2h 5.1.2h 5.1.2h		Schiphol Express Shannonweg 11 1118 LA Schiphol The Netherlands				
3 Lieu prévu par la livraison de la marchandise (lieu, pays) / Plaats (bestemd) voor de aflevering der goederen (plaats, land) / Auslieferungsort des Gutes (Ort, Land)		17 Transporteurs successifs / Opvolgende vervoerders Nachfolgende Frachtführer				
5.1.2h		344820				
4 Lieu et date de la prise en charge de la marchandise / Plaats en dat. v. inontvangstneming der goederen / Ort und Tag der Übernahme des Gutes		18 Réserves et observations du transporteur / Voorbehoud en opmerkingen van de Vervoederer / Vorbehalte und Bemerkungen des Frachtführers				
Schiphol, NL		G 5'				
5 Documents annexés / Bijgevoegde documenten Beigefügte Dokumente						
6 Marques et numéros / Merken en nummers Kennzeichen und Nummern	7 Nombre de colis / Aantal coli Anz. Packstücke	8 Mode d'emballage / Wijze van verpakking Art der Verpackung	9 Nature de la marchandise / Aard der goederen Bezeichnung des Gutes	10 No. statistique / Statistisch nummer Statistiknummer	11 Poids brut / Bruto gewicht Brutogewicht	12 Cubage / Volume Umfang
20060006 / 784-14061202	500 CT	Face Mask			3.996,0	35,500
PALLETS RUILEN + HET MISSENDE AANTAL VAN AFGELOPEN WEEK.						
25 EURDS GELOST						
EURDS GELODEN						
124						
5927						
Total: 500 3.996,0 35,500						
13 Instructions de l'expéditeur / Instructies afzender Anweisungen des Absenders		19 Conventions particulières / Speciale overeenkomsten Besondere Vereinbarungen				
LOSREF: LCH2020-0200-07L02		16 JUN '20 13:30				
14 Prescriptions d'affranchissement / Frankeringsvoorschrift / Frachtzahlungsanweisungen		20 A payer par / Te betalen door / Zu zahlen von:				
X Franco / Frei Non franco / Niet franco / Unfrei		Expéditeur / Afzender Absender				
21 Etablie à / Opgemaakt te Ausgefertigt in		Monnaie / Geldsort Währung				
Schiphol, NL		Destinataire / Geadresseerde Empfänger				
le / de am		Prix de transport / Vracht- prijs / Fracht. Reductions / Kontingen Ermäßigungen: Solde / Saldo Zwischensumme Suppléments / Supplémenten Zuschläge Frais accessoires / Bijkommen de kosten / Nebengebühren: TOTAL / TOTAAL GESAMTSUMME				
Jun 15, 2020		15 Remboursement / Rückerstattung				
22 NTG Air & Ocean Netherlands BV Shannonweg 13 1118 LA Schiphol The Netherlands		24 Marchandises reçues / Goederen ontvangen Gut empfangen				
Signature et timbre de l'expéditeur / Handtekening en stempel van de afzender / Unterschrift und Stempel des Absenders		5.1.2h				
23 Schiphol Express Shannonweg 11 1118 LA Schiphol The Netherlands						
Signature et timbre du transporteur / Handtekening en stempel van de vervoerder / Unterschrift und Stempel des Frachtführers						

1678145

1678145

1678145