



康百龄

## 广州康龄医疗用品有限公司

广东省广州市从化区明珠工业园明珠大道南 30 号 (明珠管委会斜对面高楼)  
电话: +86-20-87866686

### Internal quality inspection report 内部质检报告

sample name 样品名称	Protective Face Mask	Model 型号	KN95
Inspector 质检员		Batch/Lot No. 生产批号	200428
Packing 包装	20 / box (2pcs*10bags)	Sample Quantity 样品数量	4
Test Location 检测地点	Guangzhou 广州	Test Date 测试日期	06.05.2020
Standard 标准	GB2626-2006	Test Item 测试项目	Filtration efficiency 过滤效率

Attachment of testing result

附件: 测试结果

Sample No.  
样品号

KN95 Protective Face Mask

Sample quantity  
样品数量Test Result  
测试结果

Sample 1 样品1

Test Items 项目		
flow 流量	85.9L / Min	
Differential pressure 压差	95.0Pa	
0.3um	效率 efficiency	98.2226%
0.5um	效率 efficiency	98.7634%
1.0um	效率 efficiency	99.8398%
2.5um	效率 efficiency	100.0000%
5.0um	效率 efficiency	100.0000%
10um	效率 efficiency	100.0000%

Sample 2 样品2

Test Items 项目		
flow 流量	85.1L / Min	
Differential pressure 压差	120.5Pa	
0.3um	效率 efficiency	98.1606%
0.5um	效率 efficiency	98.6688%
1.0um	效率 efficiency	99.7554%
2.5um	效率 efficiency	100.0000%
5.0um	效率 efficiency	100.0000%
10um	效率 efficiency	100.0000%

Sample 3 样品3		
Test Items 项目	DEHS	
flow 流量	85.1L / Min	
Differential pressure 压差	156.0Pa	
0.3um	效率 efficiency	99.4686%
0.5um	效率 efficiency	99.7003%
1.0um	效率 efficiency	100.0000%
2.5um	效率 efficiency	100.0000%
5.0um	效率 efficiency	100.0000%
10um	效率 efficiency	100.0000%

Sample 4 样品4		
Test Items 项目	DEHS	
flow 流量	85.5L / Min	
Differential pressure 压差	143.0Pa	
0.3um	效率 efficiency	99.2143%
0.5um	效率 efficiency	99.4959%
1.0um	效率 efficiency	99.9643%
2.5um	效率 efficiency	100.0000%
5.0um	效率 efficiency	100.0000%
10um	效率 efficiency	100.0000%

Sample map 样品图

